



Be Right™



Sodium thiosulfate (stabilized) digital titrator cartridge, 2.26 N

Товар #: 2686901
RUB Цена: Свяжитесь с Хах
Доступно

For determination of Hypochlorite (Bleach) by Digital Titrator titration. Hach Method 10100. Range: 5 - 15% Cl₂ (50 - 150 g/L Cl₂). Stabilized sodium thiosulfate formulation. 2.26 N Digits x 0.5 = g/L Cl₂ in 0.2 mL sample

Accurate results - when and where you need them

Digital Titrator methods use a high-precision dispensation device and cartridges of concentrated titrant to allow analyses in the lab, the plant or the field.

Fast and Flexible

Interchangeable cartridges allow you to change titrants in seconds. No cleaning or rinsing burets.

Versatile

More than 50 parameter/range combinations are available

Economical

Concentrated titrants allow 40-100 typical titrations per cartridge.

Спецификации

Диапазон измерений:	50 - 150 g/L Cl ₂
Количество тестов:	40 - 100
Наименование метода:	Iodometric Method
Параметр:	Hypochlorite
Платформа:	Digital Titrator
Титрант:	2.26 N Na ₂ S ₂ O ₃